

Research Article

Linear Phase Sharp Transition BPF to Detect Noninvasive Maternal and Fetal Heart Rate

Niyan Marchon ¹, Gourish Naik ² and K. R. Pai¹

¹Padre Conceicao College of Engineering, Goa, India

²Goa University, Goa, India

Correspondence should be addressed to Niyan Marchon; niyanmarchon@gmail.com

Received 21 July 2017; Accepted 21 January 2018; Published 29 March 2018

Academic Editor: John S. Katsanis

Copyright © 2018 Niyan Marchon et al. This is an open access article distributed under the Creative Commons Attribution License, which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

Fetal heart rate (FHR) detection can be monitored using either direct fetal scalp electrode recording (invasive) or by indirect noninvasive technique. Weeks before delivery, the invasive method poses a risk factor to the fetus, while the latter provides accurate fetal ECG (FECG) information which can help diagnose fetal's well-being. Our technique employs variable order linear phase sharp transition (LPST) FIR band-pass filter which shows improved stopband attenuation at higher filter orders. The fetal frequency fiducial edges form the band edges of the filter characterized by varying amounts of overlap of maternal ECG (MECG) spectrum. The one with the minimum maternal spectrum overlap was found to be optimum with no power line interference and maximum fetal heart beats being detected. The improved filtering is reflected in the enhancement of the performance of the fetal QRS detector (FQRS). The improvement has also occurred in fetal heart rate obtained using our algorithm which is in close agreement with the true reference (i.e., invasive fetal scalp ECG). The performance parameters of the FQRS detector such as sensitivity (Se), positive predictive value (PPV), and accuracy (F_1) were found to improve even for lower filter order. The same technique was extended to evaluate maternal QRS detector (MQRS) and found to yield satisfactory maternal heart rate (MHR) results.

1. Introduction

All over the world, approximately 2.65 million stillbirths occur during pregnancy or labour especially in developing countries giving rise to the need for effective monitoring techniques with regard to fetal health [1]. FHR monitoring is important to recognize pathologic conditions, typically asphyxia, with sufficient warning so as to enable intervention by the clinician [2]. It is a screening modulus of the fetus to detect problems in advance that could result in irreversible neurological damage, even fetal death [3]. More than 85 percent of all live births in the United States undergo electronic fetal monitoring [4]. Indeed, fetal health monitoring has a significant importance in obstetrical procedures and is now widely accepted as the need of the hour.

With electronic fetal monitoring (EFM), the following expectations came: provision of accurate FECG information,

information of value in diagnosing fetal distress, prevention of fetal death or morbidity, and superiority over many methods. The fetus can be monitored electronically by two methods: direct and indirect. In the direct invasive method, the FHR is measured by a scalp electrode which is attached to the fetal scalp by means of a coiled electrode [5]. In the indirect electronic monitoring method, such as using ultrasound Doppler principle with uterine contractions, FHR can be monitored but not as precisely as the direct invasive FECG [2]. However, the invasive procedure has a risk of infection to the fetus. The ultrasound transducer with the coupling gel is applied to the mother's abdomen where fetal heart response is best detected. During this measurement, the pulsations of the maternal aorta could be detected and erroneously considered as FHR [6]. The noninvasive FECG (NIFECG) by indirect method can therefore be used to overcome all these limitations by placing the surface electrodes such as the 12 lead ECG electrodes over the maternal

abdomen [7]. The maternal thoracic ECG can also be taken as a reference signal along with the abdominal ECG (aECG). A study was conducted during labour of about 75 pregnant mothers, to check the accuracy and reliability of the NIFECG [8]. It was found that the NIFECG recordings were more accurate than the conventional external methods in comparison with the direct fetal scalp recordings. Therefore with EFM, using NIFECG recordings is the most suitable for long-term ambulatory use [9].

1.1. Fetal Physiology. Fetal distress and fetal asphyxia are too broad and vague to be applied with any precision to clinical situations. Uncertainty regarding the diagnosis based on interpretation of FHR patterns has given rise to reassuring and nonreassuring patterns [6]. Reassuring FHR patterns include the normal baseline FHR, moderate accelerations, and variability with fetal movement assuring the well-being of the fetus, whereas nonreassuring FHR patterns include tachycardia (FHR baseline more than 160 bpm), bradycardia (FHR baseline is less than 110 bpm) [10], prolonged decelerations, and so on. The severe and prolonged hypoxia induces a prolonged fall in FHR [11]. Some causes of fetal bradycardia include congenital heart block [12]. The baseline fetal heart rate is greater than 160 bpm. It is reported that the fetal body movements affect variability [13], while the baseline variability increases with advancing gestation [14]. It is also reported that reduced variability with decelerations is associated with fetal acidemia [15]. An FHR who has a consistently flat baseline with no variability and without decelerations within the normal baseline rate limit range may reflect a neurological damage in the fetus [16]. It is important that we understand the parameters of the fetal ECG signal which further aids the analysis of the fetal status and the EFM, during pregnancy or labour [1].

1.2. Previous Methods to Extract NIFECG from aECG. Researchers in the biomedical field in the areas of fetal extraction and fetal analysis have done extensive work in the last two decades. A large number of detection and extraction techniques are used to separate the FECG from the maternal ECG. Independent Component Analysis (ICA) is a statistical technique, and its accuracy is based on using a large number of noise-free maternal abdominal input channels. For ICA to function correctly, certain conditions such as (i) the number of measured signals should be equal to or greater than the number of input sources, (ii) it should possess an instantaneous linear time invariant mixing matrix, and (iii) the input sources should be statistically independent [17]. In our application, the first two do not fully satisfy because the artifacts increase the number of sources and fetal movement leads to a noninvariant mixing matrix [18]. ANFIS is an adaptive noise cancellation system which requires an additional maternal thoracic ECG signal as reference signal for adaptive cancellation of the maternal ECG. This method depends on how well one trains the ANFIS structure to compute the estimated output FECG signal [19, 20]. Subtraction method is a simple technique, but the major challenge is that the amplitude of the thoracic

MECG rarely matches the scale of the MECG present in the aECG signal [21]. As a result, correct FECG is hardly ever obtained. Wavelet transform method can be used for preprocessing stage to suppress noise, and maternal cancellation can be done by template subtraction [22]. Correlation techniques are not very efficient and effective in the detection of nonstationary signals like ECG [23]. As IIR filtering being a nonlinear method [24], our technique of using linear phase sharp transition FIR filter is less complex and does not involve many iterations as the filter response is specified precisely over the entire band. With the knowledge of the fiduciary edges and the fair estimate of the spectral overlap of maternal and fetal ECG, accurate FHR and maternal heart rate can be obtained. Our technique being single channel lead makes it very convenient and comfortable for a maternal home care for long-term monitoring.

The amplitude of MECG is at least 10 times larger than that of FECG, and the signal-to-noise ratio (SNR) of the MECG is less than unity [25]. The separation of these two ECG signals becomes even more complex as the maternal and fetal ECG overlap both in time and frequency domain [3]. The aECG signal is further affected with the low frequency noise of 0.5 Hz [26] due to baseline wandering where the amplitude of the ECG signal also varies by about 15% with respiration [2]. The other noise which affect the aECG are 50/60 Hz power line interference (PLI) [27], electromyographic noise in the uterus and the muscles of the abdomen, and other motion artefacts [28].

1.3. Linear Phase Sharp Transition Band-Pass Filters. The location of the passband and its width are critical factors that affect the design implementation of the filter. Usually sharp transition BPF are realised by the composite filters of high-pass and low-pass filters as done in [29]. The interpolated FIR technique was used wherein, every time the centre frequency of the BPF was changed, the two composite filters had to be redesigned. Approximate expressions for the value of interpolating factor and filter hardware required are derived which minimizes the total arithmetic hardware used which is the overall band-pass realization. The two-branch structure realization is more efficient than the conventional direct form realization with an increase in the number of delays [30]. Another technique of symmetric BPF is given in [31]. The filter is implemented by two parallel, quadrature filter branches with each branch derived from a complex modulation of a low-pass-interpolated FIR filter by complex exponentials. The input signal is modulated with a sine/cosine sequence in order to achieve the desired frequency shift in the frequency response.

In the current work, we propose a two-stage method to obtain noninvasive FQRS from a single lead maternal abdominal signal by first applying the designated fiduciary edges to the linear phase sharp transition (LPST) FIR band-pass filter with a sharp transition width. In the second stage, an FQRS detector is used based on Pan Tomkins QRS detector algorithm [32]. The QRS detector module consists of an amplitude squarer, moving window integrator, moving average filter, and an adaptive threshold process which effectively detects the fetal R-peaks.

2. Methodology Using LPST FIR Band-Pass Filter

Our proposed technique of integrated LPST FIR band-pass filter has low computational complexity. Normally, the composite low-pass and high-pass filters are to be redesigned for any change in the centre frequency and pass band width for the desired BPF. Our proposed technique for the integrated BPF design departs from this approach completely. It eliminates the need for a centre frequency and the fixed passband width as it is used in [33]. Our design of LPST FIR BPF allows the user to set the cutoff frequencies for a narrow pass band width. It also incorporates a very linear sharp transition width while reducing the effects due to Gibb's phenomenon and thereby reducing the passband ripple of the filter [34].

2.1. LPST FIR BPF Model and Design. In this section, the design of a LPST FIR BPF is presented. For the proposed filter model, the five regions of the filter response are modelled using trigonometric functions of frequency. The filter model magnitude response $H(\omega)$ is shown in Figure 1.

The frequency responses for the five regions are listed as follows:

$$\left. \begin{aligned} \text{Region 1 : } H(\omega) &= -\frac{\delta_s}{2} \cos(k_1 \omega) \quad 0 \leq \omega \leq \omega_{s1}, \\ \text{Region 2 : } H(\omega) &= k_2(\omega - \omega_{s1}) \quad \omega_{s1} \leq \omega \leq \omega_{p1}, \\ \text{Region 3 : } H(\omega) &= 1 + \frac{\delta_p}{2} \sin(k_3(\omega - \omega_{p1})) \quad \omega_{p1} \leq \omega \leq \omega_{p2}, \\ \text{Region 4 : } H(\omega) &= 1 - k_4(\omega - \omega_{p2}) \quad \omega_{p2} \leq \omega \leq \omega_{s2}, \\ \text{Region 5 : } H(\omega) &= -\frac{\delta_s}{2} \sin(k_5(\omega - \omega_{s2})) \quad \omega_{s2} \leq \omega \leq \pi. \end{aligned} \right\} \quad (1)$$

Using (1), the filter design parameters $k_1, k_2, k_3, k_4,$ and k_5 for the five regions of the band-pass filter are evaluated and listed as follows:

$$\left. \begin{aligned} k_1 &= \frac{2\pi m_1 + (\pi/2)}{\omega_{s1}}, \\ k_2 &= \frac{1}{(\omega_{p1} - \omega_{s1})}, \\ k_3 &= \frac{(2m_3 + 1)\pi}{(\omega_{p2} - \omega_{p1})}, \\ k_4 &= \frac{1}{(\omega_{s2} - \omega_{p2})}, \\ k_5 &= \frac{2\pi m_5 + (\pi/2)}{(\pi - \omega_{s2})}, \end{aligned} \right\} \quad (2)$$

where ω_{s1} and ω_{s2} are the stopband edge frequencies while ω_{p1} and ω_{p2} are the passband edge frequencies. δ_s and δ_p are the stopband attenuation and passband ripple, respectively, while $m_1, m_3,$ and m_5 are integers.

The impulse response coefficients $h(n)$ for the FIR band-pass filter are obtained from [35]

$$h(n) = \frac{1}{\pi} \left[\int_0^\pi H(\omega) \sin(k\omega) d\omega \right]. \quad (3)$$

Substituting the magnitude response $H(\omega)$ for each region from (1) and (2) in (3), we get

$$\begin{aligned} h(n) &= \left\{ \left(\frac{\delta_s}{4\pi} \right) \left[\frac{\cos((k+k_1)\omega_{s1}) - 1}{(k+k_1)} + \frac{\cos((k-k_1)\omega_{s1}) - 1}{(k-k_1)} \right] \right\} \\ &+ \left\{ \left(\frac{k_2}{k\pi} \right) [(-\omega_{p1}) \cos(k\omega_{p1}) + (\omega_{s1}) \cos(k\omega_{s1})] \right. \\ &- \left(\frac{k_2}{k^2\pi} \right) [\sin(k\omega_{p1}) - \sin(k\omega_{s1})] \\ &+ \left(\frac{k_2\omega_{s1}}{k\pi} \right) [\cos(k\omega_{p1}) - \cos(k\omega_{s1})] \\ &+ \left(-\frac{1}{\pi} \right) \left[\frac{\cos(k\omega_{p2}) - \cos(k\omega_{p1})}{k} \right] \\ &+ \left(\frac{\delta_p}{4\pi(k-k_3)} \right) [\sin[(k-k_3)\omega_{p2} + k_3\omega_{p1}] \\ &- \sin(k\omega_{p1})] + \left(\frac{-\delta_p}{4\pi(k+k_3)} \right) [\sin[(k+k_3)\omega_{p2} - k_3\omega_{p1}] \\ &- \sin(k\omega_{p1})] + \left(\frac{1}{k\pi} \right) [-\cos(k\omega_{s2}) + \cos(k\omega_{p2})] \\ &+ \left(\frac{k_4}{k\pi} \right) [(\omega_{s2}) \cos(k\omega_{s2}) - (\omega_{p2}) \cos(k\omega_{p2})] \\ &+ \left(\frac{k_4}{k^2\pi} \right) [\sin(k\omega_{s2}) - \sin(k\omega_{p2})] \\ &+ \left(\frac{k_4\omega_{p2}}{k\pi} \right) [-\cos(k\omega_{s2}) + \cos(k\omega_{p2})] \\ &+ \left\{ \left(\frac{-\delta_s}{4\pi} \right) \left[\left(\frac{\sin((k_5-k)\pi - k_5\omega_{s2}) + \sin(k\omega_{s2})}{(k_5-k)} \right) \right. \right. \\ &- \left. \left. \left(\frac{\sin((k_5+k)\pi - k_5\omega_{s2}) - \sin(k\omega_{s2})}{(k_5+k)} \right) \right] \right\}, \end{aligned}$$

where $k = \left[\left(\frac{N-1}{2} \right) - n \right].$

(4)

Equation (4) is the expression for the band-pass filter model impulse response $h(n)$. We can choose the effective pass band width ($\omega_{p2} \sim \omega_{p1}$) such that $(\omega_{s1} \sim \omega_{p1}) = (\omega_{s2} \sim \omega_{p2})$, as small as possible for sharp transition of passband edge. Once $\omega_{p1}, \omega_{p2}, \omega_{s1},$ and ω_{s2} are chosen, $k_1, k_2, k_3, k_4,$ and k_5 are determined.

2.2. Expression for Frequency Response Coefficients of a LPST FIR Filter. Let $h(n)$ given by (4) be the impulse response coefficients of an N point linear phase FIR filter [36] where $0 \leq n \leq N-1$ and




FIGURE 1: Magnitude response $H(\omega)$ of the band-pass filter.

$$k = \left[\left(\frac{N-1}{2} \right) - n \right], \quad n = 0, 1, 2, \dots, \left(\frac{N-3}{2} \right), \text{ for } N \text{ odd} \quad (5)$$

and

$$k = \left[\left(\frac{N-1}{2} \right) - n \right], \quad n = 0, 1, 2, \dots, \left(\frac{N}{2} - 1 \right), \text{ for } N \text{ even.} \quad (6)$$

In the case of antisymmetric response with N odd [37], the frequency response of the FIR band-pass filter is given as

$$Hr(\omega) = 2 \sum_{n=0}^{(N-3)/2} h(n) \sin\left(\omega \left(\frac{N-1}{2} - n \right)\right). \quad (7)$$

This response is most suitable for the proposed band-pass filter as $H(0) = 0$ and $H(\pi) = 0$. If we refer to (5), k is an integer for N odd. Other constraints are as follows: (i) In (2), $k \neq k_1$, $k \neq k_3$, and $k \neq k_5$ and (ii) k_1 , k_3 , and k_5 should not be integers. However, k_2 and k_4 do not have any constraints.

2.3. Fetal Frequency Spectrum. In our experiment, to extract the QRS of the MEGG and FECG from online Physionet databases [38], we used (i) Abdominal and Direct Fetal Electrocardiogram Database (adfecgdb) which provides abdominal ECG recordings (channels 2 to 5) for 5 minutes each from five different subjects during the 38–41-week gestation period [39, 40]. In addition, for each subject, a simultaneously recorded scalp or direct fetal ECG record (channel 1) is a golden reference in the evaluations to be made on the respective records. (ii) The Non-Invasive Fetal Electrocardiogram Database (nifecgdb) provides 55 records of different lengths from a single subject

taken from the 20th week of pregnancy [41]. Channels 1 and 2 represent maternal thoracic ECG signals while channels 3 to 6 are abdominal ECG recordings with only MQRS reference annotations. The Q-R-S fiducial edges of the thoracic MQRS and the invasive FQRS signals were obtained for each record. The fast Fourier transform (FFT) was obtained for the above records, an average frequency range for MQRS was found to be 10–34 Hz while the average FQRS spectrum was 20–56 Hz.

FHR varies with gestation age, ranges from 70 beats per minute (bpm) at four weeks to 175 bpm at 12 weeks and further decreases to a range of 110 to 160 bpm at full term [42]. The FECG bandwidth ranges from 0.05 to 100 Hz [2] with an average value of 140 bpm. However, in comparison, the maternal bpm normally ranges from 50 to 210 bpm with an average of 80 or 89 bpm [42].

We assumed the maternal beats per minute range to be 70–100 bpm (1.166_{\min} – 1.666_{\max} bps) and the fetal beats per minute range to be 110–140 bpm (1.833_{\min} – 2.333_{\max} bps). The minimum and maximum fetal-to-maternal (f/m) frequency ratios are obtained to compute the average f/m frequency ratio from (8) and (9).

$$\begin{aligned} \text{Minimum } \frac{f}{m} \text{ frequency ratio} &= \left(\frac{\text{fetal}_{\text{bps}}}{\text{maternal}_{\text{bps}}} \right)_{\min} \\ &= \frac{1.833}{1.166} = 1.572, \end{aligned} \quad (8)$$

$$\begin{aligned} \text{Maximum } \frac{f}{m} \text{ frequency ratio} &= \left(\frac{\text{fetal}_{\text{bps}}}{\text{maternal}_{\text{bps}}} \right)_{\max} \\ &= \frac{2.333}{1.666} = 1.400, \end{aligned} \quad (9)$$

$$\text{Average } \frac{f}{m} \text{ frequency ratio} = \left(\frac{1.572 + 1.4}{2} \right) = 1.486. \quad (10)$$

We selected the frequency spectrum for MQRS to be 18

TABLE 1: Band-pass LPST filter specifications of passband and stopband edges (type 4 fiduciary edges) along with measured magnitude response values.

Band-pass LPST filter (filter order (N) = 1001)	Stopband edge (ω_{s1}) rad/s	Passband edge (ω_{p1}) rad/s	Passband edge (ω_{p2}) rad/s	Stopband edge (ω_{s2}) rad/s
Design specifications	70π	72π	96π	98π
Measured specifications	64.68π	73.22π	92.3π	100.08π

TABLE 2: Band-pass LPST filter specifications of transition bandwidth, passband ripple, and stopband attenuation using type 4 fiduciary edges along with measured magnitude response values.

LPST filter (filter order (N) = 1001)	Transition bandwidth ($\omega_{p1}-\omega_{s1}$) rad/s	Transition bandwidth ($\omega_{s2}-\omega_{p2}$) rad/s	Max. passband loss (dB)	Min. stopband attenuation (dB)
Design specifications	2π	2π	± 0.873	40
Measured specifications	8.54π	7.78π	+0.47, -0.13	40

to 35 Hz [42] to estimate the lower and higher fetal QRS fiduciary edges from (10) as

$$\begin{aligned}
 \text{FQRS}_{\text{lower fiduciary edge}} &= \text{MQRS}_{\text{lower fiduciary edge}} \\
 &\quad \times \text{average } \frac{f}{m} \text{ frequency ratio} \\
 &= 18 \times 1.486 \sim 27 \text{ Hz}, \\
 \text{FQRS}_{\text{upper fiduciary edge}} &= \text{MQRS}_{\text{upper fiduciary edge}} \\
 &\quad \times \text{average } \frac{f}{m} \text{ frequency ratio} \\
 &= 32 \times 1.486 \sim 48 \text{ Hz}.
 \end{aligned} \tag{11}$$

From (11), we get the lower fiduciary edge of FQRS to be of 27 Hz which will remove all the low frequency noise including baseline wander frequencies and upper fiduciary edge of 48 Hz which will remove the 50 Hz and its PLI harmonics along with the high frequency noise [27]. The fetal QRS frequency band spectrum can also be further narrowed down so as to avoid the frequency band overlap of MEEG and FEKG. Accordingly, the upper fiduciary edge of the fetal QRS spectrum is chosen to be 49 Hz or 98π rad/sec. The lower fiduciary edge of the fetal spectrum is set to 35 Hz or 70π rad/sec, since the upper MQRS edge is reported to be approximately 35 Hz [42]. This results in a fetal pass band width of 14 Hz or 28π rad/sec.

2.4. FQRS Detector. To obtain the FQRS from the band-pass filtered signal, we tried looking at various algorithms including the peak-finding logic using the Hilbert transform [43]. We proposed a simple QRS detection algorithm which is based on the Pan Tomkins algorithm [32]. The modified FQRS detector comprises of four stages: (i) amplitude squarer, (ii) moving window integrator, (iii) moving average filter, and (iv) adaptive threshold. The filtered FEKG signal from the LPST FIR BPF is given to the amplitude squaring stage wherein the signal is squared point by point. This non-linear process enables the high frequency fetal R-peak signals to be further enlarged and minimizes the other lower




FIGURE 2: Illustration of the true reference FHR (direct scalp ECG) plotted with our algorithm computed FHRV for record r08 of adfecgdb (channel four) for one-minute trace. Blue dotted lines indicate ± 10 bpm tolerance with respect to the reference FHR trace.

frequency components. Further, we used a moving window integrator with a sampling frequency (f_s) of 1 KHz. This integrator effectively summed the area under the squared waveform over a fixed window interval, advanced one sample interval at a time. The width of the moving window was set to 75 sample interval for FQRS detection while a window of 152 samples wide was adjusted for MQRS detection. A too large window can merge the QRS-integrated waveform and T wave, whereas if the window is too narrow, a QRS complex could produce several peaks at the output stage [32]. Additionally, a moving average filter was also used which smoothed the integrated signal and compute a single fetal R-peak. Based on the algorithm in [44], an adaptive threshold is automatically generated to adjust to float above the unwanted




FIGURE 3: (a) Magnitude response of the proposed BPF with filter order $N = 1001$. (b) Linear plot. (c) Magnified view of the passband. (d) Magnitude response of the BPF LPST for various filter orders (N).

noise peaks. Initially, the signal peak value is adjusted manually as per the amplitude of each record [44]. The fetal R-R interval (Δn) is calculated as $(n_{i+1} - n_i)$ where n_i is the time index corresponding to the i th computed fetal R-peak at the output of the FQRS detector ($i = 1, 2, \dots$ where i is an integer). The FHR is computed for each record using

$$\text{FHR}(\text{bpm}) = \frac{(f_s \times 60)}{\Delta n}. \quad (12)$$

3. Results

3.1. Synthesis of the LPST FIR Band-Pass Filters. The LPST FIR band-pass filter was implemented using (7). The following FQRS band-pass fiduciary edge cutoff frequencies (rad/sec) were substituted as per Figure 1: $\omega_{s1} = 70\pi$, $\omega_{p1} = 72\pi$,

$\omega_{p2} = 96\pi$, and $\omega_{s2} = 98\pi$. Also stop band and passband ripple $\delta_s = \delta_p = 0.01$. Equal transition width at both ends was chosen for the pass band to be 2π rad/sec or 1 Hz. The measurement of the magnitude response of the band-pass filters is compared in Tables 1 and 2 along with the filter design specifications.

3.2. Performance Analysis of the FQRS Detector. As per the guidelines of ANSI/AAMI (ANSI/AAMI/ISO EC57 1998/ (R) 2008) [1, 45], the following classical statistics for evaluating QRS detectors were used to evaluate the FQRS detector. Sensitivity (Se), positive predictive value (PPV), and accuracy (F_1) are shown in (13) where TP, FN, and FP are true positive (correctly identified fetal R-peaks), false negative (missed fetal R-peaks), and false positive (falsely identified R-peaks), respectively. The test points assumed here are to be within

TABLE 3: Variations of passband loss and stopband attenuation for BPF with various filter orders (N).

Filter order (N)	201	501	1001	1501	2001	5001
Passband loss (dB)	1.5	± 0.5	± 0.13	± 0.1	± 0.04	± 0.03
Stopband attenuation (dB)	23.5	35.8	40.6	43	46	46




FIGURE 4: Average BPF transition width and stopband attenuation for various filter order (N).




FIGURE 5: Sensitivity, positive predictive value, and accuracy of record r08 (adfcgdb) with $TP = 132$ for various filter order (N).

± 10 bpm of their corresponding reference measurement. The true reference, namely, scalp fetal R-peak annotations from each record of the Physionet database, was compared with our experimental measured values which was implemented using Matlab toolbox. For example, we evaluated our algorithm for the adfcgdb database for the one-minute record (r08) of channel 4 and found the $TP = 132$, $FN = 1$, and $FP = 0$. The sensitivity, PPV, and F_1 were obtained to be

99.24%, 100%, and 99.61%, respectively. The average FHR values for the true reference and algorithm FHR were computed to be 132.09 bpm and 132.59 bpm, respectively. Figure 2 illustrates the true reference FHR bpm plotted with our algorithm-based fetal heart rate variability (FHRV) for record r08. The dotted lines indicate the ± 10 bpm tolerance assumed in our case with respect to the true reference trace. It was seen that the difference between the reference




FIGURE 6: (a) Direct fetal scalp ECG signal (channel one) as standard reference FECG. (b) Raw maternal aECG of adfecgdb database (channel 4—record r08).




FIGURE 7: (a) Frequency spectrum of the narrow sharp transition band-pass filtered signal. (b) FQRS time domain signal after band-pass filtering.

FHR and algorithm FHR for most R-peaks was less than the ± 8 bpm.

$$\begin{aligned}
 \text{Se} &= \frac{\text{TP}}{\text{TP} + \text{FN}}, \\
 \text{PPV} &= \frac{\text{TP}}{\text{TP} + \text{FP}}, \\
 F_1 &= 2 \left(\frac{\text{Se} \cdot \text{PPV}}{\text{Se} + \text{PPV}} \right).
 \end{aligned} \tag{13}$$

4. Discussions

We designed a LPST FIR band-pass filter such that the magnitude $H(\omega)$ in the passband and stopband are not constant but inserted a small amount of ripple of 0.01 in the stopband as well as passband so that Paley-Wiener criterion is not violated [46]. The FIR filter was designed for sharp transition width ($\omega_s - \omega_c$) of 1 Hz or 2π rad/sec. The magnitude responses of the proposed band-pass filter are shown in Figures 3(a)–3(c). Table 3 depicts the performance of the filter for various filter orders (N). There is a reduction of Gibb's




FIGURE 8: Noninvasive FHR detector for $N = 5001$ for adfecgdb database (channel 4—record r08). (a) Amplitude squaring of fetal R-peaks. (b) Moving window integration and adaptive threshold. (c) Fetal heart rate variability.

phenomenon with these filter designs. For conventional FIR sharp transition filters, the peak passband ripple due to Gibb's phenomenon is about 18% [34, 46]. In our proposed LPST FIR band-pass filters, the passband losses are quite low as can be seen from Table 2. It can also be seen that the stopband attenuation surpasses the design specification at higher orders and the passband ripple decreases for higher filter order as seen from Table 3. The sampling rate $N = 1001$ is much higher than the Nyquist rate (approximately 200 Hz) and is chosen to improve the quality of the extracted FECG. Various filter orders ($N = 201, 501, 1001, 2001,$ and 5001) were implemented to check the performance of the filters as shown in Figure 3(d). These filters are unlike the classical filters in that they possess a narrow stopband and/or passband and also sharp transition regions. The magnitude response, the linear plot, and the magnified view of the BPF are shown in Figures 3(a)–3(c), respectively, with the filter order N equal to 1001.

As seen from Figure 4, the average transition width approaches the design specifications at higher orders. The performance curves of Se, PPV, and F_1 are highly linear in the range of filter orders (N) from 2001 to 5001 as seen in Figure 5. This improvement may be due to better filtering at higher order.

The direct fetal scalp ECG is the standard reference FECG signal (channel one) as shown in Figure 6(a). The

raw maternal aECG signals were taken from channel 4—record r08 of the adfecgdb database as shown in Figure 6(b). The frequency spectrum of the signal which passed through BPF filters (frequencies between 35 Hz and 49 Hz) is shown in Figure 7(a). The band-pass filtering effectively gives us the required frequency spectrum of the FECG, which can be seen in the time domain plot in Figure 7(b).

When FQRS signal is passed through an amplitude squarer, the predefined positive peaks are prominently amplified as shown in Figure 8(a). Figure 8(b) shows the moving window integrator which integrates this signal with a selected window size, effectively picking the correct fetal R-peak indices. An illustration from Figure 8(b) shows that the time indices (n) for the first two detected fetal R-peaks are 3155 and 2709, respectively, which are above the adaptive threshold value. As shown in Figure 8(c), the FHR at these $n_{i=1}$ and $n_{i=2}$ are computed to be 134.52 bpm using (12).

Among the four types of fetal frequency fiduciary edges of the BPF, type 1 (27 Hz–53 Hz) will absorb some of the PLI in the ECG record, whereas type 2 (27 Hz–48 Hz) avoids PLI unlike type 1 but has a partial overlap spectrum of maternal ECG. Similarly, type 3 (35 Hz–53 Hz) will again have PLI problem but has no maternal spectrum overlap. Finally, the type 4 (35 Hz to 48 Hz) can be considered optimum since the maternal spectrum overlap and PLI are absent. In spite of narrowing the spectrum in this case, there are no missing




FIGURE 9: Illustration of the true reference FHR plotted with our algorithm computed FHRV for four sets of fetal frequency fiduciary edges of the BPF. The signal used is a one-minute trace of record r08, channel 4 of adfecgdb with filter order $N = 5001$ (n_i is the time index corresponding to the i th computed fetal R-peak at the output of the FQRS detector). (a) Type 1: 27 Hz–53 Hz. (b) Type 2: 27 Hz–48 Hz. (c) Type 3: 35 Hz–53 Hz. (d) Type 4: 35 Hz–48 Hz. The dotted lines indicate the ± 10 bpm tolerance assumed with respect to the reference FHR trace.




FIGURE 10: Sensitivity, positive predictive value, and accuracy for all the four channels of the 5 adfecgdb records.




FIGURE 11: Using LPST BPF for MHR detection with fiduciary edges, $\omega_{s1} = 10\pi$ and $\omega_{s2} = 40\pi$ for record r01 (channel 3) of adfcgdb. (a) Frequency spectrum of the narrow BPF signal. (b) Amplitude squaring of maternal R-peaks. (c) Moving window integration and adaptive threshold. (d) Histogram of the MHR. (e) The true reference MHR plotted with our algorithm computed MHR.




FIGURE 12: Percentage error of MHR detection compared with the MHR references for all the four channels of the 55 nifecgdb records.

fetal beats. The illustration of the true reference FHR plotted with our algorithm computed FHR for the four sets of fetal frequency fiduciary edges of the BPF is shown in Figure 9.

The FQRS detection performance parameters, Se, PPV, and F_1 , were calculated for all the four channels for each of the 5 adfecgdb records using the type 4 fetal frequency fiduciary edges as shown in Figure 10. It was observed that all the above three parameters were 100% for channel 4 of records r01 and r08. The missed fetal R-peaks (FN) were seen in some channels of records r04, r07, and r10, while the falsely identified fetal R-peaks (FP) were the least in most of the records.

It is found that this technique can be extended to detect maternal heart rate merely by changing the fiduciary edges of the BPF to $\omega_{s1}=10\pi$ and $\omega_{s2}=40\pi$ as shown in Figure 11. An illustration of the adfecgdb record r01 (channel 3) detected TP = 89, FN = 3, and FP = 0 to compute Se, PPV, and F_1 to be 96.74%, 100%, and 98.34%, respectively, as shown in Figure 11(e). Similarly, the QRS detection algorithm was tested for the MHR using the Physionet nifecgdb database for all 55 records with 3 to 4 channels each. It was observed that the MHR for all the four aECG channels for most records closely matched the MQRS reference annotations. Abdominal signals from channels 5 and 6 of records such as ecgca416, ecgca597, ecgca649, ecgca771, ecgca848, and ecgca986 displayed a large percentage error difference of computed MHR bpm value as compared with the reference MHR due to the degradation of the acquired aECG signals as seen in Figure 12.

5. Conclusion

In this paper, we described a technique of fetal heart rate detection performed noninvasively. This technique was implemented using a linear phase sharp transition FIR band-pass filter. We considered four types of fetal frequency fiduciary edges characterized by varying amounts of overlap with maternal ECG spectrum. Type 4 was found to be optimum with no PLI, no maternal spectrum overlap, and no fetal beats missed. It is found that increasing the filter order has improved the average transition bandwidth, passband ripple, and stop band attenuation of the filter. The fetal R-peaks generated by our algorithm were compared with the

scalp fetal R-peak annotations from the Physionet databases. The algorithm-generated fetal R-peaks were found to be in close agreement with each other including the average FHR values of the true reference and algorithm FHR. Similarly, other performance indices such as Se, PPV, and F_1 were found to have promising results, even for lower filter orders. The same technique was successfully extended to maternal heart rate detection.

Conflicts of Interest

The authors declare that they have no conflicts of interest.

References

- [1] J. Behar, F. Andreotti, S. Zauneder, J. Oster, and G. D. Clifford, "A practical guide to non-invasive foetal electrocardiogram extraction and analysis," *Physiological Measurement*, vol. 37, no. 5, pp. R1–R35, 2016.
- [2] M. A. Hasan, M. B. Reaz, M. I. Ibrahimy, M. S. Hussain, and J. Uddin, "Detection and processing techniques of FECG signal for fetal monitoring," *Biological Procedures Online*, vol. 11, no. 1, pp. 263–295, 2009.
- [3] M. A. Hasan, M. I. Ibrahimy, and M. B. Reaz, "Techniques of FECG signal analysis: detection and processing for fetal monitoring," *WIT Transactions on Biomedicine and Health*, vol. 12, pp. 295–305, 2007.
- [4] C. V. Ananth, S. P. Chauhan, H. Y. Chen, M. E. D'Alton, and A. M. Vintzileos, "Electronic fetal monitoring in the United States: temporal trends and adverse perinatal outcomes," *Obstetrics & Gynecology*, vol. 121, no. 5, pp. 927–933, 2013.
- [5] R. Sameni and G. D. Clifford, "A review of fetal ECG signal processing issues and promising directions," *The Open Pacing, Electrophysiology & Therapy Journal*, vol. 3, pp. 4–20, 2010.
- [6] M. M. Corton, K. J. Leveno, S. L. Bloom, C. Y. Spong, and J. S. Dashe, *Williams Obstetrics 24/E*, McGraw Hill Professional, 2014.
- [7] G. D. Clifford, I. Silva, J. Behar, and G. B. Moody, "Non-invasive fetal ECG analysis," *Physiological Measurement*, vol. 35, no. 8, pp. 1521–1536, 2014.
- [8] W. R. Cohen, S. Ommani, S. Hassan et al., "Accuracy and reliability of fetal heart rate monitoring using maternal abdominal surface electrodes," *Acta Obstetrica et Gynecologica Scandinavica*, vol. 91, no. 11, pp. 1306–1313, 2012.

- [9] M. Peters, J. Crowe, J. F. Piéri et al., "Monitoring the fetal heart non-invasively: a review of methods," *Journal of Perinatal Medicine*, vol. 29, no. 5, pp. 408–416, 2001.
- [10] N. Manassiev, "What is the normal heart rate of a term fetus?," *BJOG: An International Journal of Obstetrics & Gynaecology*, vol. 103, no. 12, pp. 1272–1273, 1996.
- [11] A. S. Thakor and D. A. Giussani, "Effects of acute acidemia on the fetal cardiovascular defense to acute hypoxemia," *American Journal of Physiology-Regulatory, Integrative and Comparative Physiology*, vol. 296, no. 1, pp. R90–R99, 2009.
- [12] E. T. Jaeggi and M. K. Friedberg, "Diagnosis and management of fetal bradyarrhythmias," *Pacing and Clinical Electrophysiology*, vol. 31, no. s1, pp. S50–S53, 2008.
- [13] H. P. van Geijn, H. W. Jongsma, J. de Haan, and T. K. Eskes, "Analysis of heart rate and beat-to-beat variability: interval difference index," *American Journal of Obstetrics & Gynecology*, vol. 138, no. 3, pp. 246–252, 1980.
- [14] M. Pillai and D. James, "The development of fetal heart rate patterns during normal pregnancy," *Obstetrics & Gynecology*, vol. 76, no. 5, pp. 812–816, 1990.
- [15] R. H. Paul, A. K. Suidan, S. Yeh, B. S. Schifrin, and E. H. Hon, "Clinical fetal monitoring: VII. The evaluation and significance of intrapartum baseline FHR variability," *American Journal of Obstetrics & Gynecology*, vol. 123, no. 2, pp. 206–210, 1975.
- [16] R. K. Freeman, T. J. Garite, M. P. Nageotte, and L. A. Miller, *Fetal heart rate monitoring*, Lippincott Williams & Wilkins, 2012.
- [17] Y. Ye, Z. L. Zhang, J. Zeng, and L. Peng, "A fast and adaptive ICA algorithm with its application to fetal electrocardiogram extraction," *Applied Mathematics and Computation*, vol. 205, no. 2, pp. 799–806, 2008.
- [18] M. Varanini, G. Tartarisco, L. Billeci, A. Macerata, G. Pioggia, and R. Balocchi, "An efficient unsupervised fetal QRS complex detection from abdominal maternal ECG," *Physiological Measurement*, vol. 35, no. 8, pp. 1607–1619, 2014.
- [19] C. K. Vijila, P. Kanagasabapathy, and S. Johnson, "Adaptive neuro fuzzy inference system for extraction of FECG," in *INDICON, 2005 Annual IEEE*, Chennai, India, December 2005.
- [20] J.-S. R. Jang, "ANFIS: adaptive-network-based fuzzy inference system," *IEEE Transactions on Systems, Man, and Cybernetics*, vol. 23, no. 3, pp. 665–685, 1993.
- [21] P. Bergveld and W. J. Meijer, "A new technique for the suppression of the MECG," *IEEE Transactions on Biomedical Engineering*, vol. BME-28, no. 4, pp. 348–354, 1981.
- [22] J. C. Echeverria, N. Ramirez, A. B. Pimentel, R. Rodriguez, R. Gonzalez, and V. Medina, "Fetal QRS extraction based on wavelet analysis and pattern matching," in *Proceedings of the 18th Annual International Conference of the IEEE Engineering in Medicine and Biology Society, 1996. Bridging Disciplines for Biomedicine*, Amsterdam, Netherlands, November 1996.
- [23] J. H. van Bommel, "Detection of weak foetal electrocardiograms by autocorrelation and crosscorrelation of envelopes," *IEEE Transactions on Biomedical Engineering*, vol. BME-15, no. 1, pp. 17–23, 1968.
- [24] A. Kam and A. Cohen, "Detection of fetal ECG with IIR adaptive filtering and genetic algorithms," in *1999 Proceedings, 1999 IEEE International Conference on Acoustics, Speech, and Signal Processing*, Phoenix, AZ, USA, March 1999.
- [25] D. J. Jagannath and A. I. Selvakumar, "Issues and research on foetal electrocardiogram signal elicitation," *Biomedical Signal Processing and Control*, vol. 10, pp. 224–244, 2014.
- [26] G. Lenis, N. Pilia, A. Loewe, W. H. Schulze, and O. Dössel, "Comparison of baseline wander removal techniques considering the preservation of ST changes in the ischemic ECG: a simulation study," *Computational and Mathematical Methods in Medicine*, vol. 2017, Article ID 9295029, 13 pages, 2017.
- [27] D. D. Taralunga, I. Gussi, and R. Strungaru, "Fetal ECG enhancement: adaptive power line interference cancellation based on Hilbert Huang Transform," *Biomedical Signal Processing and Control*, vol. 19, pp. 77–84, 2015.
- [28] E. C. Karvounis, M. G. Tsiouras, D. I. Fotiadis, and K. K. Naka, "An automated methodology for fetal heart rate extraction from the abdominal electrocardiogram," *IEEE Transactions on Information Technology in Biomedicine*, vol. 11, no. 6, pp. 628–638, 2007.
- [29] Y. Neuvo, G. Rajan, and S. A. Mitra, "Design of narrow-band FIR bandpass digital filters with reduced arithmetic complexity," *IEEE Transactions on Circuits and Systems*, vol. 34, no. 4, pp. 409–419, 1987.
- [30] G. Rajan, Y. Neuvo, and S. K. Mitra, "On the design of sharp cutoff wide-band FIR filters with reduced arithmetic complexity," *IEEE Transactions on Circuits and Systems*, vol. 35, no. 11, pp. 1447–1454, 1988.
- [31] F. Mintzer and B. Liu, "Practical design rules for optimum FIR bandpass digital filters," *IEEE Transactions on Acoustics, Speech, and Signal Processing*, vol. 27, no. 2, pp. 204–206, 1979.
- [32] J. Pan and W. J. Tompkins, "A real-time QRS detection algorithm," *IEEE Transactions on Biomedical Engineering*, vol. BME-32, no. 3, pp. 230–236, 1985.
- [33] J. X. Rodrigues and K. R. Pai, "Modified linear phase frequency response masking FIR filter," in *Proceedings of the 4th International Symposium on Image and Signal Processing and Analysis, 2005. ISPA 2005*, Zagreb, Croatia, September 2005.
- [34] J. R. Johnson, *Introduction to Digital Signal Processing*, Prentice Hall, Englewood Cliffs, NJ, 1989.
- [35] R. N. Bracewell, *The Fourier Transform and Its Applications*, vol. 31999, McGraw-Hill, New York, 1986.
- [36] L. Rabiner, J. Kaiser, and R. Schafer, "Some considerations in the design of multiband finite-impulse-response digital filters," *IEEE Transactions on Acoustics, Speech, and Signal Processing*, vol. 22, no. 6, pp. 462–472, 1974.
- [37] J. G. Proakis and D. G. Manolakis, *Digital Signal Processing: Principles Algorithms and Applications*, 2001.
- [38] A. L. Goldberger, L. A. Amaral, L. Glass et al., "PhysioBank, PhysioToolkit, and PhysioNet: components of a new research resource for complex physiologic signals," *Circulation*, vol. 101, no. 23, pp. e215–e220, 2000.
- [39] J. Jezewski, A. Matonia, T. Kupka, D. Roj, and R. Czabanski, "Determination of fetal heart rate from abdominal signals: evaluation of beat-to-beat accuracy in relation to the direct fetal electrocardiogram," *Biomedical Engineering*, vol. 57, no. 5, pp. 383–394, 2012.
- [40] "Abdominal and direct fetal electrocardiogram database (adfecgdb)," Available: <https://www.physionet.org/physiobank/database/adfecgdb/>.
- [41] "Non-invasive fetal electrocardiogram database (nifecgdb)," Available: <https://physionet.org/physiobank/database/nifecgdb/>.
- [42] G. Lamesgin, Y. Kassaw, and D. Assefa, "Extraction of fetal ECG from abdominal ECG and heart rate variability analysis,"

- in *Afro-European Conference for Industrial Advancement*, Springer International Publishing, 2015.
- [43] M. S. Manikandan and K. P. Soman, "A novel method for detecting R-peaks in electrocardiogram (ECG) signal," *Biomedical Signal Processing and Control*, vol. 7, no. 2, pp. 118–128, 2012.
- [44] N. Marchon and G. Naik, "QRS detector for maternal abdominal ECG," in *International Conference on Signal and Information Processing (IconSIP)*, Vishnupuri, India, October 2016.
- [45] ANSI/AAMI/ISO, "Testing and reporting performance results of cardiac rhythm and ST-segment measurement algorithms," 2008, EC57 1998/(R).
- [46] R. E. Paley and N. Wiener, *Fourier Transforms in the Complex Domain*, vol. 19, American Mathematical Society, New York, 1934.



Hindawi

Submit your manuscripts at
<https://www.hindawi.com>

